# Product Data Sheet

Frank Dental GmbH, Toelzer Str. 4, 83703 Gmund, Germany
Fon: +49-8022-6736-0, Fax: +49-8022-6736-900
Email: info@frank-dental.de
IFU under: https://www.frank-dental.com
Opening hours: Mon - Thu: 8-18, Fri: 8-13

frank.dental
technologie vom tegernsee

**Article description**
D.807.016.C.FG

**UMDNS**
16-670

**ISO standard**
806.314.225.504.016

**VPE**
5

**Dimensions (length)**
19



## Information on components/composition
Art: Diamant bur, Stahl Schaft + Kopf diamantiert
Color:

## Possible hazards
With proper handling (consideration of handling and speed, face protection and face mask etc.), no danger, water hazard class: 0 (self-classification)

## First aid measures
no special measures required

## Fire-fighting measures
Product does not burn. Ambient fire with appropriate extinguishing agents in accordance with fire protection regulations.

## Measures in case of accidental release
not applicable

## Handling and storage
Use in contra-angle handpiece according to recommended speed, store dry, avoid contamination after cleaning by autoclaves, temperature fluctuations and contact with acids, avoid chemicals

- sterilize instrument before use 

- Speed limitation 225.000
- Consider working angle and contact pressure
- Check tight fit in drive element
- Observe safety and operating instructions

## Exposure controls and personal protective equipment
General instructions for use and safety



## Physical and chemical properties
Odour: none, Flash point/density/viscosity: not applicable, Solubility in water: insoluble

## Stability and reactivity
no dangerous reactions when handled properly

## Information on toxicology
non-toxic

## Ecological information
no harmful effect known

## Note on disposal
Waste key number: 35 101 - Disposal with household waste in solid container permitted

## Transport information
no dangerous goods

## 14. Directives
certified according to EN ISO 13485:2016 + AC:2018 + A11:2021 and medical device directive 93/42/EEC
water hazard class: 0 (self-assessment)

## 15. Intended purpose



## Further information
The user is responsible for proper use in accordance with the intended purpose and meets hygiene requirements and for ensuring that reprocessing with the materials, equipment, etc. used is successful. Regular checks of the procedures described here are therefore required and any deviations must be documented by the user.

## Further information
Frank Dental always recommends wearing safety glasses and a face mask when using rotating instruments both for the protection of the personnel in the practice, as well as for the protection of the patient. Usage: Cavity preparation, finishing of fillings, chamfer preparation, etc.
Existing laws and regulations on the use of such products must be observed by the user.

Release by responsible person: Armin Frank - Status: January 2025



CE 0297